





中国认可 国际互认 检测 TESTING CNAS L0599

Test Report SL52025258316901TX Date: July 03,2020 Page 1 of 9

BAODING ZHUOSHU MEDICAL DEVICE MANUFACTURING CO., LTD.

RENLI VILLAGE, ANSU TOWN, XUSHUI DISTRICT, BAODING CITY.

The following sample(s) was/were submitted and identified on behalf of the client as:

Sample Description : (A)Filtering half mask

Sample Color : (A)white Model No. : ZS2099

Manufacturer : Baoding Zhuoshu medical device manufacturing Co., Ltd.

Country of Origin : China Country of Destination : EUR

Specifications : 14cmx10cm

Sample Receiving Date : May 22, 2020

Testing Period : May 22, 2020 – Jul 03, 2020

Test Result(s) : Unless otherwise stated the results shown in this test report refer only to the

sample(s) tested, for further details, please refer to the following page(s).


Test Performed : Selected test(s) as requested by applicant

Conclusion:

Sample No.	Recommendation Level
(A)	FFP1 NR

Signed for and on behalf of

SGS-CSTC Standards Technical Services (Shanghai) Co., Ltd Testing Center





Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at http://www.sgs.com/en/Terms-and-Conditions.apx and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Document.aspx.
Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the Ilmits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate the except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this lest report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention: To check the authenticity of testing /inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or, email: CN.Doccheck@sgs.com